

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.473 °C	30.1 °C	-0.373 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.576 %HR	25.5 %HR	-3.076 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.908 %HR	36.1 %HR	-14.808 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.824 %HR	52.1 %HR	-17.724 %HR	± 5.0 %HR	± 1.9E+00 %HR