

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.473 °C	30.2 °C	-0.273 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.576 %HR	32.6 %HR	4.024 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.908 %HR	51.1 %HR	0.192 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.824 %HR	68.6 %HR	-1.224 %HR	± 5.0 %HR	± 1.9E+00 %HR