

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.473 °C	30.5 °C	0.027 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.576 %HR	27.0 %HR	-1.576 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.908 %HR	39.3 %HR	-11.608 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.824 %HR	54.2 %HR	-15.624 %HR	± 5.0 %HR	± 1.9E+00 %HR