

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.544 °C	30.9 °C	0.356 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.801 %HR	31.1 %HR	2.299 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.272 %HR	51.3 %HR	2.028 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	71.685 %HR	70.4 %HR	-1.285 %HR	± 3.0 %HR	± 1.9E+00 %HR