

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.544 °C	30.8 °C	0.256 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.801 %HR	31.2 %HR	2.399 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.272 %HR	51.2 %HR	1.928 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	71.685 %HR	68.9 %HR	-2.785 %HR	± 3.0 %HR	± 1.9E+00 %HR