

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 10 °C                      | 9.997 °C                   | 10.0 °C            | 0.003 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 20 °C                      | 20.155 °C                  | 19.9 °C            | -0.255 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 °C                      | 30.544 °C                  | 30.7 °C            | 0.156 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 40 °C                      | 40.451 °C                  | 40.2 °C            | -0.251 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |