

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	9.997 °C	9.9 °C	-0.097 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.155 °C	19.9 °C	-0.255 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.544 °C	30.4 °C	-0.144 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.451 °C	40.1 °C	-0.351 °C	± 0.5 °C	± 4.0E-01 °C