

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.972 °C	29.6 °C	-0.372 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.801 %HR	32.0 %HR	3.199 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.621 %HR	53.0 %HR	2.379 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.353 %HR	79.0 %HR	4.647 %HR	± 5.0 %HR	± 1.9E+00 %HR