

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.211 °C	15 °C	-0.211 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.473 °C	30 °C	-0.473 °C	± 1.0 °C	± 4.0E-01 °C
50 %HR	50.908 %HR	51 %HR	0.092 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.824 %HR	71 %HR	1.176 %HR	± 5.0 %HR	± 1.9E+00 %HR