

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
220.518 µl	220 µl	-0.518 µl	± 1.6 µl	±	5.2E-02 µl
999.86 µl	1000 µl	0.135 µl	± 8.0 µl	±	5.9E-02 µl
2,000.41 µl	2000 µl	-0.411 µl	± 16 µl	±	9.4E-02 µl