

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
10.033 µl	10 µl	-0.033 µl	± 0.12 µl	±	5.1E-02 µl
50.07 µl	50 µl	-0.065 µl	± 0.50 µl	±	5.1E-02 µl
100.33 µl	100 µl	-0.331 µl	± 0.80 µl	±	5.2E-02 µl