

Patrón <i>(Standard)</i>		UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
10 Hz	0.80 g-pk	0.81 g-pk	0.01 g-pk	± 0.04 g-pk	±	1.7E-02 g-pk
30 Hz	1.00 g-pk	1.00 g-pk	0.00 g-pk	± 0.05 g-pk	±	2.1E-02 g-pk
50 Hz	1.00 g-pk	1.01 g-pk	0.01 g-pk	± 0.05 g-pk	±	2.1E-02 g-pk
100 Hz	1.00 g-pk	1.01 g-pk	0.01 g-pk	± 0.05 g-pk	±	2.1E-02 g-pk
500 Hz	1.00 g-pk	1.01 g-pk	0.01 g-pk	± 0.05 g-pk	±	2.1E-02 g-pk
1000 Hz	1.00 g-pk	1.02 g-pk	0.02 g-pk	± 0.05 g-pk	±	2.2E-02 g-pk