

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.303 °C	30.41 °C	0.107 °C	± 0.5 °C	± 4.0E-01 °C
65 %HR	64.854 %HR	66 %HR	1.146 %HR	± 3.0 %HR	± 1.9E+00 %HR