

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.353 °C	30.26 °C	-0.093 °C	± 0.5 °C	± 4.0E-01 °C
65 %HR	65.059 %HR	67 %HR	1.941 %HR	± 3.0 %HR	± 1.9E+00 %HR