

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.323 °C	30.44 °C	0.117 °C	± 0.5 °C	± 4.0E-01 °C
65 %HR	64.527 %HR	67 %HR	2.473 %HR	± 3.0 %HR	± 1.9E+00 %HR