

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.198 °C	10.8 °C	0.602 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.045 °C	20.1 °C	0.055 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.473 °C	29.9 °C	-0.573 °C	± 1.0 °C	± 4.0E-01 °C
50 %HR	50.908 %HR	48 %HR	-2.908 %HR	± 5.0 %HR	± 1.9E+00 %HR