

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.75	0.75173	-1.730	± 1.8	± 2.9E+00 mm					
0 a 0.8	0.80173	-1.730	± 1.8	± 2.9E+00 mm					
0 a 0.9	0.90172	-1.720	± 1.9	± 2.9E+00 mm					
0 a 1	1.00172	-1.720	± 2	± 2.9E+00 mm					
0 a 1.2	1.20271	-2.710	± 2.1	± 2.9E+00 mm					
0 a 1.4	1.40270	-2.700	± 2.3	± 2.9E+00 mm					
0 a 1.6	1.60269	-2.690	± 2.4	± 2.9E+00 mm					
0 a 1.7	1.70269	-2.690	± 2.5	± 2.9E+00 mm					
0 a 1.8	1.80268	-2.680	± 2.6	± 2.9E+00 mm					
0 a 1.9	1.90267	-2.670	± 2.7	± 2.9E+00 mm					