

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.932 °C	30.5 °C	0.568 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.333 %HR	30.4 %HR	1.067 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.026 %HR	49.3 %HR	0.274 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.039 %HR	74.3 %HR	-0.739 %HR	± 3.0 %HR	± 1.9E+00 %HR