

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.342 °C	15.5 °C	0.158 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.520 °C	25.4 °C	-0.120 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.604 °C	30.4 °C	-0.204 °C	± 1.0 °C	± 4.0E-01 °C
60 %HR	59.936 %HR	65.3 %HR	5.364 %HR	± 3.5 %HR	± 1.9E+00 %HR
65 %HR	64.670 %HR	69.5 %HR	4.830 %HR	± 3.5 %HR	± 1.9E+00 %HR
70 %HR	70.366 %HR	74.5 %HR	4.134 %HR	± 3.5 %HR	± 1.9E+00 %HR