

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 15 °C             | 15.342 °C         | 15.5 °C   | 0.158 °C      | ± 1.0 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.520 °C         | 24.8 °C   | -0.720 °C     | ± 1.0 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.604 °C         | 29.9 °C   | -0.704 °C     | ± 1.0 °C             | ± 4.0E-01 °C                |
| 60 %HR            | 59.936 %HR        | 63.1 %HR  | 3.164 %HR     | ± 3.5 %HR            | ± 1.9E+00 %HR               |
| 65 %HR            | 64.670 %HR        | 66.7 %HR  | 2.030 %HR     | ± 3.5 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 70.366 %HR        | 71.1 %HR  | 0.734 %HR     | ± 3.5 %HR            | ± 1.9E+00 %HR               |