

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.342 °C	15.2 °C	-0.142 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.520 °C	24.8 °C	-0.720 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.604 °C	29.8 °C	-0.804 °C	± 1.0 °C	± 4.0E-01 °C
60 %HR	59.936 %HR	64.5 %HR	4.564 %HR	± 3.5 %HR	± 1.9E+00 %HR
65 %HR	64.670 %HR	68.2 %HR	3.530 %HR	± 3.5 %HR	± 1.9E+00 %HR
70 %HR	70.366 %HR	73.4 %HR	3.034 %HR	± 3.5 %HR	± 1.9E+00 %HR