

Certificado No. (Certificate #):		65117	Equipo (Instrument):			Micrómetro de Exterior
Fecha de Calibración (Calibration Date):		2026-01-16	Marca (Brand):			PEL
Patrón (Standard)		UBP (UUT)	Error (Error)	E.M.P (Tolerance)		Incertidumbre (Uncertainty)
25.00	mm	25.00 mm	0 µm	± 4 µm		± 5.8E+00 µm
27.00	mm	27.00 mm	0 µm	± 4 µm		± 5.8E+00 µm
30.00	mm	30.00 mm	0 µm	± 4 µm		± 5.8E+00 µm
33.00	mm	33.00 mm	0 µm	± 4 µm		± 5.8E+00 µm
35.00	mm	35.00 mm	0 µm	± 4 µm		± 5.8E+00 µm
38.00	mm	38.00 mm	0 µm	± 4 µm		± 5.8E+00 µm
40.00	mm	40.00 mm	0 µm	± 4 µm		± 5.8E+00 µm
42.00	mm	42.00 mm	0 µm	± 4 µm		± 5.8E+00 µm
45.00	mm	45.00 mm	0 µm	± 4 µm		± 5.8E+00 µm
47.00	mm	47.00 mm	0 µm	± 4 µm		± 5.8E+00 µm
50.00	mm	50.00 mm	0 µm	± 4 µm		± 5.8E+00 µm