

Certificado No. (Certificate #):	65121		Equipo (Instrument):	Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):	2026-01-16		Marca (Brand):	Mitutoyo	
Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Exterior / (Exterior)	5.00 mm	5.00 mm	0 µm	± 50 µm	± 1.5E+03 µm
Exterior / (Exterior)	15.00 mm	15.00 mm	0 µm	± 50 µm	± 4.5E+03 µm
Exterior / (Exterior)	30.00 mm	30.00 mm	0 µm	± 50 µm	± 9.0E+03 µm
Exterior / (Exterior)	80.00 mm	80.00 mm	0 µm	± 50 µm	± 2.4E+04 µm
Exterior / (Exterior)	150.00 mm	150.00 mm	0 µm	± 50 µm	± 4.5E+04 µm
Interior / (Interior)	45.00 mm	45.00 mm	0 µm	± 50 µm	± 1.4E+04 µm
Interior / (Interior)	130.00 mm	130.00 mm	0 µm	± 50 µm	± 1.5E+03 µm
Profundidad / (Depth)	50.00 mm	50.00 mm	0 µm	± 50 µm	± 1.5E+04 µm
Profundidad / (Depth)	100.00 mm	100.00 mm	0 µm	± 50 µm	± 3.0E+04 µm