

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	29.7 °C	-0.362 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	28.0 %HR	-2.161 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.325 %HR	49.0 %HR	-1.325 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.499 %HR	69.0 %HR	-1.499 %HR	± 5.0 %HR	± 1.9E+00 %HR