

Certificado No. (Certificate #):

65167

Equipo (Instrument):

Reloj Indicador

Fecha de Calibración (Calibration Date):

2026-01-15

Marca (Brand):

Humboldt

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1.000 mm	0.921 mm	-79 µm		±	5.8E-01 µm
5.000 mm	4.584 mm	-416 µm		±	5.8E-01 µm
10.000 mm	9.147 mm	-853 µm		±	5.8E-01 µm
14.000 mm	12.802 mm	-1198 µm		±	5.8E-01 µm
16.000 mm	14.636 mm	-1365 µm		±	5.8E-01 µm
18.000 mm	16.460 mm	-1540 µm		±	5.8E-01 µm
20.000 mm	18.279 mm	-1721 µm		±	5.8E-01 µm
22.000 mm	20.106 mm	-1894 µm		±	5.8E-01 µm
25.000 mm	22.848 mm	-2152 µm		±	5.8E-01 µm