

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	19.874 °C	20.0 °C	0.126 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.922 °C	29.8 °C	-0.122 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.919 °C	39.8 °C	-0.119 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	49.169 %HR	53.4 %HR	4.231 %HR	± 5.0 %HR	± 1.9E+00 %HR
60 %HR	59.619 %HR	55.3 %HR	-4.309 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.632 %HR	67.5 %HR	-3.112 %HR	± 5.0 %HR	± 1.9E+00 %HR