

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
20 °C	19.874 °C	19.9 °C	0.016 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.922 °C	29.7 °C	-0.222 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.919 °C	39.7 °C	-0.219 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	49.169 %HR	53.7 %HR	4.531 %HR	± 5.0 %HR	± 1.9E+00 %HR
60 %HR	59.619 %HR	62.6 %HR	2.941 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.632 %HR	69.5 %HR	-1.152 %HR	± 5.0 %HR	± 1.9E+00 %HR