

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	19.874 °C	19.7 °C	-0.174 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.922 °C	29.7 °C	-0.222 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.919 °C	39.6 °C	-0.319 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	49.169 %HR	53.6 %HR	4.431 %HR	± 5.0 %HR	± 1.9E+00 %HR
60 %HR	59.619 %HR	55.4 %HR	-4.189 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.632 %HR	66.5 %HR	-4.112 %HR	± 5.0 %HR	± 1.9E+00 %HR