

Certificado No. (Certificate #):

65178

Equipo (Instrument):

Reloj Palpador

Fecha de Calibración (Calibration Date):

2026-01-15

Marca (Brand):

Humboldt

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
0.050 in	0.05 in	0 in		±	6.1E-04 in
0.100 in	0.10 in	0 in		±	6.2E-04 in
0.105 in	0.11 in	0 in		±	6.2E-04 in
0.150 in	0.15 in	0 in		±	6.2E-04 in
0.200 in	0.20 in	0 in		±	6.3E-04 in
0.300 in	0.30 in	0 in		±	6.5E-04 in
0.400 in	0.40 in	0 in		±	6.7E-04 in
0.500 in	0.50 in	0 in		±	7.0E-04 in
1.000 in	1.00 in	0 in		±	9.3E-04 in