

Certificado No. (Certificate #):		65179	Equipo (Instrument):		Reloj Palpador
Fecha de Calibración (Calibration Date):		2026-01-15	Marca (Brand):		Humboldt
Patrón (Standard)		UBP (UUT)	Error (Error)	Exactitud (Accuracy)	
1.000 mm		0.997 mm	-4 µm	±	5.8E-01 µm
1.100 mm		1.096 mm	-4 µm	±	5.8E-01 µm
1.500 mm		1.498 mm	-2 µm	±	5.8E-01 µm
2.000 mm		1.996 mm	-4 µm	±	5.8E-01 µm
3.000 mm		2.995 mm	-5 µm	±	5.8E-01 µm
4.000 mm		3.994 mm	-6 µm	±	5.8E-01 µm
5.000 mm		4.994 mm	-6 µm	±	5.8E-01 µm
6.000 mm		5.992 mm	-8 µm	±	5.8E-01 µm
6.300 mm		6.299 mm	-1 µm	±	5.8E-01 µm