

Certificado No. (Certificate #):

65181

Equipo (Instrument):

Reloj Indicador

Fecha de Calibración (Calibration Date):

2026-01-15

Marca (Brand):

Humboldt

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1.000 mm	0.999 mm	-1 µm		±	5.8E-01 µm
1.100 mm	1.098 mm	-2 µm		±	5.8E-01 µm
1.500 mm	1.501 mm	1 µm		±	5.8E-01 µm
2.000 mm	2.000 mm	0 µm		±	5.8E-01 µm
3.000 mm	3.002 mm	2 µm		±	5.8E-01 µm
4.000 mm	4.001 mm	1 µm		±	5.8E-01 µm
5.000 mm	5.000 mm	0 µm		±	5.8E-01 µm
6.000 mm	6.000 mm	0 µm		±	5.8E-01 µm
6.300 mm	6.307 mm	7 µm		±	5.8E-01 µm