

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
199.5 RPM	200 RPM	0.5 RPM	± 5 RPM	±	5.8E+01 RPM
498 RPM	500 RPM	2.0 RPM	± 5 RPM	±	5.8E+01 RPM
996 RPM	1000 RPM	4.0 RPM	± 10 RPM	±	5.8E+01 RPM
1493 RPM	1500 RPM	7.0 RPM	± 15 RPM	±	5.8E+01 RPM
1990 RPM	2000 RPM	10.0 RPM	± 20 RPM	±	5.8E+01 RPM