

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
198.7 RPM	200 RPM	1.3 RPM	± 5 RPM	±	5.8E+01 RPM
495 RPM	500 RPM	5.0 RPM	± 5 RPM	±	5.8E+01 RPM
992 RPM	1000 RPM	8.0 RPM	± 10 RPM	±	5.8E+01 RPM
1488 RPM	1500 RPM	12.0 RPM	± 15 RPM	±	5.8E+01 RPM
1985 RPM	2000 RPM	15.0 RPM	± 20 RPM	±	5.8E+01 RPM