

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| 52.65 μm          | 53.06 μm  | 0.41 μm       | ± 1.6 μm          | ±                           | 5.8E-02 μm |
| 155.23 μm         | 155.2 μm  | -0.03 μm      | ± 4.7 μm          | ±                           | 6.2E-02 μm |
| 257.12 μm         | 256.6 μm  | -0.52 μm      | ± 7.7 μm          | ±                           | 7.0E-02 μm |
| 511.72 μm         | 511.6 μm  | -0.12 μm      | ± 15.4 μm         | ±                           | 9.7E-02 μm |