

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.792 °C	30.0 °C	0.208 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.967 %HR	34.6 %HR	4.633 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.343 %HR	52.9 %HR	3.557 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.445 %HR	72.8 %HR	3.355 %HR	± 5.0 %HR	± 1.9E+00 %HR