

Certificado No. (Certificate #):		65258		Equipo (Instrument):		Micrómetro de Interior			
Fecha de Calibración (Calibration Date):		2026-01-21		Marca (Brand):		Mitutoyo			
Patrón (Standard)		UBP (UUT)		Error (Error)		E.M.P (Tolerance)		Incertidumbre (Uncertainty)	
2.2000 in		2.1904 in		-0.0096 in		± 0.0003 in		±	1.2E-04 in
2.3000 in		2.2904 in		-0.0096 in		± 0.0003 in		±	1.2E-04 in
2.4000 in		2.3904 in		-0.0096 in		± 0.0003 in		±	1.2E-04 in
2.5000 in		2.4904 in		-0.0096 in		± 0.0003 in		±	1.2E-04 in
2.6000 in		2.5904 in		-0.0096 in		± 0.0003 in		±	1.2E-04 in