

Certificado No. (Certificate #):		65259		Equipo (Instrument):		Micrómetro de Interior			
Fecha de Calibración (Calibration Date):		2025-06-16		Marca (Brand):		Mitutoyo			
Patrón (Standard)		UBP (UUT)		Error (Error)		E.M.P (Tolerance)		Incertidumbre (Uncertainty)	
0.8000 in		0.79150 in		-0.0085 in		± 0.0003 in		±	2.9E-05 in
1.0000 in		1.00880 in		0.0088 in		± 0.0003 in		±	2.9E-05 in
1.0000 in		0.98700 in		-0.0130 in		± 0.0003 in		±	2.9E-05 in
1.2000 in		1.20500 in		0.0050 in		± 0.0003 in		±	2.9E-05 in
1.2000 in		1.19700 in		-0.0030 in		± 0.0003 in		±	2.9E-05 in
1.6000 in		1.60700 in		0.0070 in		± 0.0003 in		±	2.9E-05 in
1.6000 in		1.58900 in		-0.0110 in		± 0.0003 in		±	2.9E-05 in
2.0000 in		2.00800 in		0.0080 in		± 0.0003 in		±	2.9E-05 in