

Certificado No. (Certificate #):		65260		Equipo (Instrument):		Micrómetro de Interior			
Fecha de Calibración (Calibration Date):		2026-01-21		Marca (Brand):		Mitutoyo			
Patrón (Standard)		UBP (UUT)		Error (Error)		E.M.P (Tolerance)		Incertidumbre (Uncertainty)	
2.0000 in		2.00800 in		0.00800 in		± 0.0003 in		±	2.9E-05 in
2.5000 in		2.50460 in		0.00460 in		± 0.0003 in		±	2.9E-05 in
2.5000 in		2.50340 in		0.00340 in		± 0.0003 in		±	2.9E-05 in
3.0000 in		3.00390 in		0.00390 in		± 0.0003 in		±	2.9E-05 in
3.0000 in		2.99520 in		-0.00480 in		± 0.0003 in		±	2.9E-05 in
3.5000 in		3.50690 in		0.00690 in		± 0.0003 in		±	2.9E-05 in
3.5000 in		3.49820 in		-0.00180 in		± 0.0003 in		±	2.9E-05 in
4.0000 in		3.98260 in		-0.01740 in		± 0.0003 in		±	2.9E-05 in