

Certificado No. (Certificate #):		65261		Equipo (Instrument):		Micrómetro de Interior			
Fecha de Calibración (Calibration Date):		2026-01-21		Marca (Brand):		Mitutoyo			
Patrón (Standard)		UBP (UUT)		Error (Error)		E.M.P (Tolerance)		Incertidumbre (Uncertainty)	
4.0000 in		4.00200 in		0.00200 in		± 0.0003 in		±	2.9E-05 in
5.0000 in		5.00290 in		0.00290 in		± 0.0005 in		±	2.9E-05 in
5.0000 in		5.00220 in		0.00220 in		± 0.0005 in		±	2.9E-05 in
6.0000 in		6.00290 in		0.00290 in		± 0.0005 in		±	2.9E-05 in
6.0000 in		6.00150 in		0.00150 in		± 0.0005 in		±	2.9E-05 in
7.0000 in		7.00030 in		0.00030 in		± 0.0005 in		±	2.9E-05 in
7.0000 in		6.99970 in		-0.00030 in		± 0.0005 in		±	2.9E-05 in
8.0000 in		8.00140 in		0.00140 in		± 0.0005 in		±	2.9E-05 in