

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.208 °C	10.1 °C	-0.108 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.7 °C	0.451 °C	± 1.0 °C	± 4.0E-01 °C
40 °C	40.491 °C	41.4 °C	0.909 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	33.8 %HR	3.230 %HR	± 10.0 %HR	± 1.9E+00 %HR
50 %HR	50.100 %HR	51.2 %HR	1.100 %HR	± 10.0 %HR	± 1.9E+00 %HR
80 %HR	80.161 %HR	84.5 %HR	4.339 %HR	± 10.0 %HR	± 1.9E+00 %HR