

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 15 °C                      | 15.222 °C                  | 15.29 °C           | 0.068 °C               | ± 0.21 °C                     | ± 4.0E-01 °C                         |
| 20 °C                      | 20.436 °C                  | 20.51 °C           | 0.074 °C               | ± 0.21 °C                     | ± 4.0E-01 °C                         |
| 30 °C                      | 30.463 °C                  | 30.52 °C           | 0.057 °C               | ± 0.21 °C                     | ± 4.0E-01 °C                         |
| 30 %HR                     | 30.570 %HR                 | 35.20 %HR          | 4.630 %HR              | ± 3.50 %HR                    | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.100 %HR                 | 54.40 %HR          | 4.300 %HR              | ± 3.50 %HR                    | ± 1.9E+00 %HR                        |
| 80 %HR                     | 80.161 %HR                 | 83.50 %HR          | 3.339 %HR              | ± 3.50 %HR                    | ± 1.9E+00 %HR                        |