

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.130 °C	-25.4 °C	-0.270 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.090 °C	-0.3 °C	-0.390 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.224 °C	4.8 °C	-0.424 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.017 °C	9.8 °C	-0.217 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.299 °C	25.3 °C	0.001 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.524 °C	30.2 °C	-0.324 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.501 °C	40.5 °C	-0.001 °C	± 0.5 °C	± 4.0E-01 °C

