

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
500.062 RPM	500 RPM	-0.062 RPM	± 5 RPM	±	1.3E-01 RPM
1299.972 RPM	1300 RPM	0.028 RPM	± 13 RPM	±	1.3E-01 RPM