

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
18 °C	18.029 °C	18.06 °C	0.031 °C	± 0.21 °C	± 4.0E-01 °C
20 °C	20.436 °C	20.26 °C	-0.176 °C	± 0.21 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.06 °C	-0.189 °C	± 0.21 °C	± 4.0E-01 °C
35 °C	35.397 °C	35.59 °C	0.193 °C	± 0.21 °C	± 4.0E-01 °C
40 %HR	40.488 %HR	38.90 %HR	-1.588 %HR	± 3.50 %HR	± 1.9E+00 %HR
75 %HR	75.458 %HR	73.50 %HR	-1.958 %HR	± 3.50 %HR	± 1.9E+00 %HR
80 %HR	80.018 %HR	78.77 %HR	-1.248 %HR	± 3.50 %HR	± 1.9E+00 %HR