

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.133 °C	29.9 °C	-0.233 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.813 %HR	34 %HR	4.187 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.826 %HR	55 %HR	4.174 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.578 %HR	74 %HR	4.422 %HR	± 5.0 %HR	± 1.9E+00 %HR