

Certificado No. (Certificate #):

65434

Equipo (Instrument):

Reloj Palpador

Fecha de Calibración (Calibration Date):

2026-01-15

Marca (Brand):

Humboldt

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
0.050 in	0.046 in	-0.005 in	± in	±	1.3E-03 in
0.100 in	0.095 in	-0.005 in	± in	±	1.8E-03 in
0.150 in	0.145 in	-0.005 in	± in	±	1.8E-03 in
0.200 in	0.195 in	-0.005 in	± in	±	1.8E-03 in
0.300 in	0.294 in	-0.006 in	± in	±	8.4E-04 in
0.500 in	0.493 in	-0.007 in	± in	±	1.9E-03 in
1.000 in	0.993 in	-0.007 in	± in	±	2.0E-03 in
1.500 in	1.493 in	-0.007 in	± in	±	2.1E-03 in
1.800 in	1.796 in	-0.004 in	± in	±	1.4E-03 in