

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>		Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1799.656 s	1800 s	0.344 s	0.019%	± 0.01 %	±	1.7E+01 s
1799.778 s	1800 s	0.222 s	0.012%	± 0.01 %	±	1.7E+01 s
1799.821 s	1800 s	0.179 s	0.010%	± 0.01 %	±	1.7E+01 s