

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.593 °C	15.5 °C	-0.093 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.597 °C	24.4 °C	-0.197 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.048 °C	31.8 °C	-0.248 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.586 %HR	29.1 %HR	0.514 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	71.460 %HR	69.8 %HR	-1.660 %HR	± 3.0 %HR	± 1.9E+00 %HR