

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
Sensor Interno					
8 °C	7.570 °C	8.5 °C	0.930 °C	± 1.0 °C	± 4.0E-01 °C
12 °C	11.652 °C	12.4 °C	0.748 °C	± 1.0 °C	± 4.0E-01 °C
50 %HR	49.854 %HR	51 %HR	1.146 %HR	± 5.0 %HR	± 1.9E+00 %HR
Sensor Externo					
8 °C	7.570 °C	8.0 °C	0.430 °C	± 1.0 °C	± 4.0E-01 °C
12 °C	11.652 °C	11.9 °C	0.248 °C	± 1.0 °C	± 4.0E-01 °C
50 %HR	49.854 %HR	53 %HR	3.146 %HR	± 5.0 %HR	± 1.9E+00 %HR