

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.593 °C	15.4 °C	-0.193 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	24.798 °C	24.5 °C	-0.298 °C	± 1.0 °C	± 4.0E-01 °C
50 %HR	49.854 %HR	52 %HR	2.146 %HR	± 5.0 %HR	± 1.9E+00 %HR